

U.S. NUCLEAR REGULATORY COMMISSION**TELEPHONE CONVERSATION RECORD**Date: 8/15, 8/24, 8/27 &
8/28/2007

Time: 1400,1530,0935 & 0845

Mail Control 2007-001 License No(s). 37-10913-06
or Report No(s).

Docket No(s). 03029497

Name of Licensee: Defense Logistics Agency

Name of Participant(s): Craig Lindsey, Safety
Joseph Jaworski, Radiation Safety Officer
Steven R. Courtemanche, Health Physicist NRC RITelephone No. 215-737-3240
215-737-3246
610-337-5075

Subject: Priority 5T Inspection

(NOTE: This will be used as the
Documents Title in ADAMS)

Summary: Mr. Lindsey indicated that Mr. Jaworski would be the appropriate person to speak to about the licensed program and that he would not be available until Monday at about 8 or 9 a.m.
8/24/07 15:30 Mr. Lindsey stated that Mr. Jaworski was not in but would be back on Monday.
8/27/07 09:35 I left a message on Mr. Jaworski's voice-mail.
8/28/2007 08:45 I conducted a 5T inspection. In addition to the information on the 5T record, the following was determined:
The licensee has six ECD with three in use. The serial numbers are 3433, 3530, 3765 (all in storage) and 3795, 3495 and 3496. They are Perkin Elmer ECD. The ones in storage have not been used in about four years. A generally-licensed gauge is kept in the locked cabinet with the ECD. (Microderm) It contains a Thallium-204 sealed source and is leak tested every six months even though it has not been used in years.

Action Required: Place in ADAMS

Document Availability:	<input checked="" type="checkbox"/> Publicly Available	<input type="checkbox"/> Non-Publicly Available	
X Non-Sensitive	<input type="checkbox"/> Non-Sensitive Copyright	<input type="checkbox"/> Sensitive	<input type="checkbox"/> Sensitive Copyright
Immediate Release	<input checked="" type="checkbox"/> Normal Release	<input type="checkbox"/> Delay Release Date	
Prepared & SUNSI Review Completed By:	Steven R. Courtamanche	/ RA /	Date: 8/28/2007