

From: Michelle Beardsley
To: LAT@nrc.gov
Date: Mon, Aug 27, 2007 10:10 AM
Subject: Fwd: Medical Diagnostic Labs application

>>> Michelle Beardsley 08/27/2007 10:52 AM >>>
(License No. 29--31257-01) not issued
Docket No. 03037497
Control No. 140708

Tom, after reviewing your application dated August 10, 2007, the following additional information is needed:

1. Please indicate the manufacturer and model no. of the Cs-137 sealed source to be used for calibrations-- also indicate who will be performing instrument calibrations and their experience/education(DO NOT send lists of publications with CV);
2. Indicate the isotopes that have been used/handled by each of your proposed authorized users in the past (except for Nichols- his CV included this info)--Please DO NOT Send CV's; only the isotopes used by each;
3. Please specify if the proposed authorized user- Mr. Gallagher has a college degree;
4. You state in this application that you will be doing iodinations with I-125, please confirm that this will be in a fume hood and that air monitoring will be performed to ensure that effluent limits as stated in Part 20 Appendix B will not be exceeded; also indicate the frequency that bioassays will be done on those individuals performing these studies.

Please respond in writing; you may fax (see no. below) your response to my attention. Be sure to put the Control No. listed above on your correspondence. If you fax your response-- do not mail a copy. Please respond within 15 days or we shall assume that you do not wish to pursue this application.

Michelle

Michelle Beardsley
Senior Health Physicist
U.S.N.R.C.
Region I
475 Allendale Road
King of Prussia, PA 19406
610-337-6942
fax 610-337-5393

Mail Envelope Properties (46D2E949.E00 : 16 : 55716)

Subject: Fwd: Medical Diagnostic Labs application
Creation Date Mon, Aug 27, 2007 10:10 AM
From: Michelle Beardsley

Created By: MRB@nrc.gov

Recipients

nrc.gov
kp1_po.KP_DO
LAT (LAT@nrc.gov)

Post Office
kp1_po.KP_DO

Route
nrc.gov

Files	Size	Date & Time
MESSAGE	2629	Monday, August 27, 2007 10:10 AM

Options

Expiration Date: None
Priority: Standard
ReplyRequested: No
Return Notification:
Send Notification when Opened

Concealed Subject: No
Security: Standard

Junk Mail Handling Evaluation Results

Message is not eligible for Junk Mail handling
Message is from an internal sender

Junk Mail settings when this message was delivered

Junk Mail handling disabled by User
Junk List is not enabled
Junk Mail using personal address books is not enabled
Block List is not enabled