

<b>U.S. NUCLEAR REGULATORY COMMISSION</b>		Date: 8/28/07
<b>TELEPHONE CONVERSATION RECORD</b>		Time:
Mail Control or Report No(s).	140330	License No(s). 45-00034-30 Docket No(s). 03003067
Name of Licensee:	University of Virginia (gamma knife license)	
Name of Participant(s):	Alan Aqualino, Ph.D., RSO and AMP	
Telephone No.	484-982-0781	
Subject: <small>(NOTE: This will be used as the Documents Title in ADAMS)</small>	Review of letter dated 8/21/07	
Summary:	<p>I informed AI that nearly all information required for clinical use of Perfexion has been provided. We will await the following before issuing the amendment:</p> <ul style="list-style-type: none"> <li>a) Documentation of completion of Dr. Lerner's training</li> <li>b) Description of how licensee will address timer termination in daily spot-checks</li> <li>c) Elekta's procedure for 6-month spot-check of sector positions</li> </ul> <p>I also explained to AI that we are able to approve additional AUs and AMPs only after receiving documentation that the described Perfexion training has been completed. These should be simple amendments that can be issued rapidly.</p>	
Action Required:	Await the additional information listed above, then issue amendment	
Document Availability:	<input checked="" type="checkbox"/> Publicly Available <input type="checkbox"/> Non-Publicly Available <input checked="" type="checkbox"/> Non-Sensitive <input type="checkbox"/> Non-Sensitive Copyright <input type="checkbox"/> Sensitive <input type="checkbox"/> Sensitive Copyright <input type="checkbox"/> Immediate Release <input checked="" type="checkbox"/> Normal Release <input type="checkbox"/> Delay Release Date	
Prepared & SUNSI Review Completed By:	S. Gabriel	/ RA /      Date: 8/29/07