

ATTN: Toy Siemens

**Troxler Electronic Laboratories, Inc.**

3008 Cornwallis Rd., P.O. Box 12057

Research Triangle Park, NC 27709

Tel: (877) 876-9537 Fax: (919) 485-2250

License: NC 032-0182-1

MIRERJ MELETA  
ENGINEERING SERVICES, INC  
32300 SCHOOLCRAFT ROAD  
SUITE D-1  
LIVONIA, MI 48150

---

**LEAK TEST CERTIFICATE**


---

**DEVICE:****Model:** 3430**Serial No:** 26415**SEALED SOURCES:**

Serial No.	Measure Date	Nuclide	GBq	mCi
75-9533	05/08/1996	CS-137	0.296	8
47-23001	06/05/1996	AM-241:BE	1.48	40

**LEAK TEST ANALYSIS:****Sample collected on:** 04/24/2007**Sample analyzed on:** 05/04/2007 at 10:47:00 AM**Analyzed by:** HARRY P. PENDLETON, JR.

	ALPHA	BETA-GAMMA
Conversion factor (cpm/Bq)	1.29E+01	2.02E+01
Background measurement (cpm)	1	25
Sample measurement (cpm)	0	23
Activity (Bq)	< MDA	< MDA
Min. Detectable Activity (Bq)	5.1E-01	1.3E+00

**This certifies that the leak test results are:** **Less than 185 Bq (0.005 uCi)** **Greater than 185 Bq (0.005 uCi)**