



UNITED STATES
NUCLEAR REGULATORY COMMISSION
REGION I
475 ALLENDALE ROAD
KING OF PRUSSIA, PENNSYLVANIA 19406-1415

June 3, 2008

Docket No. 030-36502
Control No. 142423

License No. 45-30881-01

Daniel G. Amen, M.D.
Director
Amen Clinics, Inc.
1875 Campus Commons Drive
Suite 101
Reston, VA 20191

**SUBJECT: AMEN CLINICS, INC., REQUEST FOR ADDITIONAL INFORMATION
CONCERNING APPLICATION FOR AMENDMENT TO LICENSE, CONTROL
NO. 142423**

Dear Dr. Amen:

This is in reference to your letter dated May 27, 2008 requesting to amend Nuclear Regulatory Commission License No. 45-30881-01. In order to continue our review, we need the following additional information:

In support of your request to authorize Dr. Amen as the Radiation Safety Officer, in addition to the information provided (the copy of the license on which the physician was specifically named as an authorized user for the uses requested, and the description of the experience in radiation safety procedures), you must also provide:

- a. a brief description of the training in radiation safety, regulatory issues and emergency procedures for the medical uses authorized on your license, using the table in Part I, Item 3.c. of Form 313A (RSO) that is enclosed with this letter. This table also requires that you list the name of the supervising individual, who must be a radiation safety officer, authorized user, authorized nuclear pharmacist, or authorized medical physicist for those medical uses and list the license/permit number of the supervising individual.
- b. a preceptor attestation, using one of the options in Part II of Form 313A(RSO) enclosed with this letter. The preceptor does not have to be the supervising individual as long as the preceptor provides, directs, or verifies that the above training and experience has been satisfactorily completed and that a level of competency sufficient to function independently as an authorized user for the medical uses authorized has been achieved.
- f. If applicable, description of recent (within the past 7 years) related continuing education and experience as required by 10 CFR 35.59 if the board certification, or training and experience, was received more than 7 years ago.

D. Amen
Amen Clinics, Inc.

2

Current NRC regulations and guidance are included on the NRC's website at www.nrc.gov; select **Nuclear Materials; Medical, Academic, and Industrial Uses of Nuclear Material**; then **Regulations, Guidance, and Communications**. You may also obtain these documents by contacting the Government Printing Office (GPO) toll-free at 1-866-512-1800. The GPO is open from 7:00 a.m. to 8:00 p.m. EST, Monday through Friday (except Federal holidays).

We will continue our review upon receipt of this information. Please reply to my attention at the Region I Office and refer to Mail Control No. 142423. If you have any technical questions regarding this deficiency letter, please call me at (610) 337-5040.

If we do not receive a reply from you within 30 calendar days from the date of this letter, we will assume that you do not wish to pursue your application.

Sincerely,

Original signed by Elizabeth Ullrich

Betsy Ullrich
Senior Health Physicist
Commercial and R&D Branch
Division of Nuclear Materials Safety

Enclosure:
Form 313A(RSO)

DOCUMENT NAME: C:\FileNet\ML081560401.wpd

SUNSI Review Complete: EUIrich

After declaring this document "An Official Agency Record" it will be released to the Public.

To receive a copy of this document, indicate in the box: "C" = Copy w/o attach/encl "E" = Copy w/ attach/encl "N" = No copy

OFFICE	DNMS/RI	<input checked="" type="checkbox"/> N	DNMS/RI	<input type="checkbox"/>	DNMS/RI	<input type="checkbox"/>	<input type="checkbox"/>
NAME	EUIrich / <i>EAU</i> /						
DATE	6/3/08						

OFFICIAL RECORD COPY