

From: Nguyen, Janice
Sent: Friday, January 13, 2012 1:26 PM
To: 'rbrouillette@lmhosp.org'
Cc: 'radrem@aol.com'
Subject: NRC Amendment Request for Lawrence & Memorial Hospital (Mail Control Number 576422)

Licensee: Lawrence & Memorial Hospital
License Number: 06-09261-01
Docket Number: 030-01275
Mail Control: 576422

Dear Mr. Brouillette,

This is regarding the amendment request for Lawrence & Memorial Hospital dated November 18, 2011 to add an outpatient PET/CT facility at 196 Parkway South, Suite 102, Waterford, Connecticut. The following additional information is needed to allow us to continue our review of your request:

1. Please confirm that 196 Parkway South, Suite 102, Waterford, Connecticut, is not a location of use for any other active NRC license.
2. Please confirm if there is any lead in the hot lab door.
3. Please identify any additional authorized users involved with your PET/CT program and provide their training and experience as described in 10 CFR 35.290. NRC Form 313A (AUD) may be used to document this information.
4. You indicated you would be using the Medrad Intego PET Infusion system. Please confirm the following:
 - a. Will the dose calibrator included with the unit be the one you will use to assay patient dosages? If so, please provide what quality assurance testing will be performed on the unit.
 - b. Please provide the shielding for the device and the maximum activity you intend to use with it.

You may reply to my attention by mail to the Region I office, by fax to 610-337-5269, or by return email if the response has a signed letter from management and is in a pdf format. Please reference mail control 576422 in your response.

Could you please reply to this email to confirm receipt?

Thank you for your cooperation. Please feel free to contact me by telephone or e-mail if you have any questions.

Sincerely,

Jan

Janice Nguyen
Health Physicist
US Nuclear Regulatory Commission Region I
475 Allendale Road
King of Prussia, PA 19406
(610) 337-5006
FAX (610) 337-5269