

**Internal
Medicine
Consultants II
L.L.C.**



MATTHEW S. BOSNER, M.D., F.A.C.C. ■ *CARDIOVASCULAR DISEASES*
FRANCES T. MCKINNEY, M.D., PhD. ■ *INTERNAL MEDICINE*
Suite 299 – 456 N. New Ballas Road – St. Louis, MO 63141
Telephone (314) 569-1090 – Fax (314) 569-1424

RECEIVED

JAN - 6 2012

DNMS

Jan. 4, 2012

Materials Licensing Section

U.S. Nuclear Regulatory Commission, Region IV

611 Ryan Plaza Drive, Suite 400

Arlington, TX 76011-8064

Re: RAM License # 24-32472-01

This letter is a request to decommission the restricted areas under this license.

This lab only used Tc99m except for a use of Tl-201 over a year ago during the Tc99m shortage.

This last use of Tc99m was on Dec. 30, 2011. Enclosed is the wipes and surveys and document showing the transfer of the sources to another license.


Please expedite this request so the leased areas may be quickly released back to the property owners.

Thank You,

A handwritten signature in cursive script that reads "Lauren Bean".


Lauren Bean, CNMT

Technical Director



Accredited Echocardiography
Laboratory



NRC FORM 314 <small>(7-2004)</small> <small>10 CFR 30.260(c), 40.430(f), 70.380(f), and 72.540(x)</small>	U.S. NUCLEAR REGULATORY COMMISSION	APPROVED BY OMB: NO. 3150-0028 EXPIRES: 07/31/2004 <small>Estimated burden per response to comply with this mandatory collection request: 30 minutes. This material is used by NRC as part of the basis for its determination that the facility is released for unrestricted use. Send comments regarding burden estimate to the Records Management Branch (T-6 ER), U.S. Nuclear Regulatory Commission, Washington, DC 20555-0001, or by internet e-mail to tjs1@nrc.gov and to the Desk Officer, Office of Information and Regulatory Affairs, NEOS-1120-2, (3150-0028), Office of Management and Budget, Washington, DC 20503. If a meter used to impose an information collection does not display a currently valid OMB control number, the NRC may not conduct or sponsor, and a person is not required to respond to, the information collection.</small>
CERTIFICATE OF DISPOSITION OF MATERIALS		
LICENSEE NAME AND ADDRESS Internal Medicine Consultants/HC 456 N. New Ballas Rd St. Louis, Mo. 63141	LICENSE NUMBER 24-32472-01	DOCKET NUMBER 030-36388
A. LICENSE STATUS (Check the appropriate box) <input type="checkbox"/> This license has expired. <input checked="" type="checkbox"/> This license has not yet expired, please terminate it.		
B. DISPOSAL OF RADIOACTIVE MATERIAL <small>(Check the appropriate boxes and complete as necessary. If additional space is needed, provide attachments)</small> The licensee, or any individual executing this certificate on behalf of the licensee, certifies that: <ul style="list-style-type: none"> <input type="checkbox"/> 1. No radioactive materials have ever been procured or possessed by the licensee under this license. <input checked="" type="checkbox"/> 2. All activities authorized by this license have ceased, and all radioactive materials procured and/or possessed by the licensee under this license number cited above have been disposed of in the following manner: <ul style="list-style-type: none"> <input checked="" type="checkbox"/> a. Transfer of radioactive materials to the licensee listed below: on 1/3/2012 Town & Country Cardiovascular 1010 Old Des Peres Rd St. Louis, Mo. 63131 <input type="checkbox"/> b. Disposal of radioactive materials: <ul style="list-style-type: none"> <input type="checkbox"/> 1. Directly by the licensee: <input type="checkbox"/> 2. By licensed disposal site: <input type="checkbox"/> 3. By waste contractor: <input type="checkbox"/> c. All radioactive materials have been removed such that any remaining residual radioactivity is within the limits of 10 CFR Part 20, Subpart E, and is ALARA. 		
C. SURVEYS PERFORMED AND REPORTED <ul style="list-style-type: none"> <input checked="" type="checkbox"/> 1. A radiation survey was conducted by the licensee. The survey confirms: <ul style="list-style-type: none"> <input checked="" type="checkbox"/> a. the absence of licensed radioactive materials <input type="checkbox"/> b. that any remaining residual radioactivity is within the limits of 10 CFR 20, Subpart E, and is ALARA. <input checked="" type="checkbox"/> 2. A copy of the radiation survey results: <ul style="list-style-type: none"> <input checked="" type="checkbox"/> a. is attached; or <input type="checkbox"/> b. is not attached (Provide explanation); or <input type="checkbox"/> c. was forwarded to NRC on: _____ Date <input type="checkbox"/> 3. A radiation survey is not required as only sealed sources were ever possessed under this license, and <ul style="list-style-type: none"> <input type="checkbox"/> a. The results of the latest leak test are attached; and/or <input type="checkbox"/> b. No leaking sources have ever been identified. 		
The person to be contacted regarding the information provided on this form:		
<small>NAME</small> Lauren Bean	<small>TITLE</small> Cert Nuclear Med Tech	<small>TELEPHONE (include Area Code)</small> 314-569-1090
<small>E-MAIL ADDRESS</small> laurenbean2003@yahoo.com		
<small>Mail all future correspondence regarding this license to:</small> Matthew Bosner		
C. CERTIFYING OFFICIAL I CERTIFY UNDER PENALTY OF PERJURY THAT THE FOREGOING IS TRUE AND CORRECT		
<small>PRINTED NAME AND TITLE</small> Lauren Bean CNMT	<small>SIGNATURE</small> 	<small>DATE</small> 1/3/12
<small>WARNING: FALSE STATEMENTS IN THIS CERTIFICATE MAY BE SUBJECT TO CIVIL AND/OR CRIMINAL PENALTIES. NRC REGULATIONS REQUIRE THAT SUBMISSIONS TO THE NRC BE COMPLETE AND ACCURATE IN ALL MATERIAL RESPECT. 18 U.S.C. SECTION 1001 MAKES IT A CRIMINAL OFFENSE TO MAKE A WILLFULLY FALSE STATEMENT OR REPRESENTATION TO ANY DEPARTMENT OR AGENCY OF THE UNITED STATES AS TO ANY MATTER WITHIN ITS JURISDICTION.</small>		

Town & Country Cardiovascular Group

BJC Medical Group

John P. Hess, M.D., F.A.C.C.
Stuart T. Higano, M.D., F.A.C.C., F.S.C.A.I.
Howard S. Lite, M.D., F.A.C.C., F.A.S.E.
Robert B. Lehman, M.D., F.A.C.C.
Terri L. Kolker, PA-C

1010 Old Des Peres Road
St. Louis, MO 63131
Phone: (314) 238-2535
Fax: (314) 238-2020

January 03, 2012

Lauren Bean/Dr. Bosner
Internal Medicine Consultants, LLC.
456 North New Ballas Rd.
St. Louis, MO 63141

RE: SEALED SOURCE TRANSFER

Dear Lauren/Dr. Bosner:

This written document acknowledges the permanent transfer of (2) two NRC licensed sources and one exempt source. The sources were received on 01/03/12, at Town and Country Cardiovascular Group (NRC Lic. # 24-32661-01).

The following two NRC licensed sealed sources (Part 35.65) were received:


1. Cs-137, vial, 197.4 μ Ci, 7/01/03, Isotope Products, Mod. SRV-137-200u, S/N 988-39-21.
2. Co-57, flood, 10 mCi, 02/01/11, Isotope Products, Mod. NES 8510, S/N 1474-001.

The above sources were last leak tested on 11/16/11, and removable contamination was below 0.005 μ Ci on each source.

The following exempt source was received:

1. Cs-137, rod, 101.6 nCi, 7/01/03, Isotope Products, Mod. # SGF-0014, Source # 962-44-2.

Sincerely,



Tom Dickinson
Radiation Safety Officer
Town and Country Cardiovascular Group

Location: Internal Medicine Consultants
456 N. New Ballas Rd. Suite 299 St. Louis MO 63141

Survey Area	Net results: mR/Hr	Date:	01/03	01/04
2012	Action Level	Time:	15:55	10:14
		Bkgnd:	0.02	0.02

	Survey Instrument	Lab-ID/Initials:	na/LB	na/LB
			0.1	0.1
Hot Lab Floor			0	0
			1	1
Camera Floor			0	0
			0.1	0.1
Treadmill			0	0
			0.1	0.1
Hot Lab Dose Prep Area			0	0
			1	1
Hands / Feet			0	0
			0.1	0.1

Survey Instrument List for summary reports

Lab ID	Instrument	Make	Model	Serial number	Last calibration
	Capintec CRC-15Well	Capintec	CRC 15W	170731	2/16/11
	Ludlum 14C	Ludlum	14C	197904	12/27/10
	Ludlum 14C	Ludlum	14C	197904	12/15/11

Personnel Legend

LB LAUREN BEAN

Radiation Safety Officer: _____

Internal Medicine Consultants

456 N. New Ballas Rd., Suite 299

St. Louis, MO 63141

Dose Calibrator Channel Check Report

1/3/12 to 1/4/12

Location: Internal Medicine Consultants
456 N. New Ballas Rd. Suite 299 St. Louis MO 63141

Date Time Initials	Base reading Measured activity Percent error	Instrument	Serial#: 170731 Make: Capintec	Model: CRC 15-W Source: Cs-137	Serial#: 988-39-21
--------------------------	--	-------------------	-----------------------------------	-----------------------------------	--------------------

Test date	Cs137	Tc 99m	Tl 201
1/3/12	170.45	313.64	178.98
08:42:00	168.00	310.00	178.00
LB	-1.3%	-1.1%	-0.4%
1/4/12	170.44	313.61	178.96
10:16:43	168.00	310.00	178.00
LB	-1.3%	-1.1%	-0.4%

Personnel Legend

LB LAUREN BEAN

Radiation Safety Officer: _____