

CERTIFICATE OF DISPOSITION OF MATERIALS

PLEASE READ THESE INSTRUCTIONS BEFORE COMPLETING NRC FORM 314.

Subpart E of 10 CFR Part 20 establishes the radiological criteria for license terminations/decommissioning of facilities licensed under 10 CFR Parts 30, 40, 50, 60, 61, 70, and 72, as well as other facilities subject to the Commission's jurisdiction under the Atomic Energy Act of 1954, as amended, and the Energy Reorganization Act of 1974, as amended.

INSTRUCTIONS

Section B, Item 2.

Licensees should describe the specific radioactive material transfer actions. If radioactive wastes were generated in terminating this license, the licensee should describe the disposal actions taken, including the disposition of low-level radioactive waste, mixed waste, greater-than-Class-C waste, and sealed sources.

Section B, Item 2.a.

The information provided concerning the transfer of radioactive material to another licensee should specify the date of the transfer, the name of the licensee recipient, an individual contact name and telephone number for the licensee recipient, and the recipient's NRC or Agreement State license number.

Section B, Item 2.b.

For disposal of radioactive materials, licensees should describe the specific disposal method or procedure (e.g., decay-in-storage). For those cases when radioactive materials are disposed of by a licensed disposal site or by a waste contractor, the licensee should specify the name, address, and telephone number of the licensed disposal site operator or waste contractor.

Section B, Item 2.c.

"Residual radioactivity," as defined in 10 CFR 20.1003, means radioactivity in 'areas' (structures, materials, soils, etc.) remaining as a result of activities (licensed and unlicensed) under the licensee's control from sources used by the licensee, excluding background radiation. ALARA is defined in 10 CFR 20.1003.

FILE CERTIFICATES AS FOLLOWS:

IF YOU ARE LOCATED IN:

ALABAMA, CONNECTICUT, DELAWARE, DISTRICT OF COLUMBIA, FLORIDA, GEORGIA, KENTUCKY, MAINE, MARYLAND, MASSACHUSETTS, NEW HAMPSHIRE, NEW JERSEY, NEW YORK, NORTH CAROLINA, PENNSYLVANIA, PUERTO RICO, RHODE ISLAND, SOUTH CAROLINA, TENNESSEE, VERMONT, VIRGINIA, VIRGIN ISLANDS, OR WEST VIRGINIA, SEND CERTIFICATES TO:

LICENSING ASSISTANT SECTION
NUCLEAR MATERIALS SAFETY BRANCH
U.S. NUCLEAR REGULATORY COMMISSION, REGION I
475 ALLENDALE ROAD
KING OF PRUSSIA, PA 19406-1415

IF YOU ARE LOCATED IN:

ALASKA, ARIZONA, ARKANSAS, CALIFORNIA, COLORADO, HAWAII, IDAHO, KANSAS, LOUISIANA, MISSISSIPPI, MONTANA, NEBRASKA, NEVADA, NEW MEXICO, NORTH DAKOTA, OKLAHOMA, OREGON, PACIFIC TRUST TERRITORIES, SOUTH DAKOTA, TEXAS, UTAH, WASHINGTON, OR WYOMING, SEND CERTIFICATES TO:

MATERIAL RADIATION PROTECTION SECTION
U. S. NUCLEAR REGULATORY COMMISSION, REGION IV
1600 E. LAMAR BOULEVARD
ARLINGTON, TX 76011-4511

ILLINOIS, INDIANA, IOWA, MICHIGAN, MINNESOTA, MISSOURI, OHIO, OR WISCONSIN, SEND CERTIFICATES TO:

MATERIALS LICENSING SECTION
U.S. NUCLEAR REGULATORY COMMISSION, REGION III
2443 WARRENVILLE ROAD, SUITE 210
LISLE, IL 60532-4352

RECEIVED FEB 22 2012

(1-2012)
10 CFR 30.36(j)(1); 40.42(j)(1);
70.38(j)(1); and 72.54(k)(5)(1)(1)

Estimated burden per response to comply with this mandatory collection request: 30 minutes. This submittal is used by NRC as part of the basis for its determination that the facility is released for unrestricted use. Send comments regarding burden estimate to the Information Services Branch (T-5 F53), U.S. Nuclear Regulatory Commission, Washington, DC 20555-0001, or by internet e-mail to Infocollects.Resource@nrc.gov, and to the Desk Officer, Office of Information and Regulatory Affairs, NEOB-10202, (3150-0028), Office of Management and Budget, Washington, DC 20503. If a means used to impose an information collection does not display a currently valid OMB control number, the NRC may not conduct or sponsor, and a person is not required to respond to, the information collection.

CERTIFICATE OF DISPOSITION OF MATERIALS

LICENSEE NAME AND ADDRESS

RMT, INC.
744 Heartland Trail
PO Box 8923
Madison, WI 53708-8923

LICENSE NUMBER

48-21408-01

DOCKET NUMBER

0302-0678

LICENSE EXPIRATION DATE

09/30/2012

A. LICENSE STATUS (Check the appropriate box)

- This license has expired. This license has not yet expired; please terminate it.

B. DISPOSAL OF RADIOACTIVE MATERIAL

(Check the appropriate boxes and complete as necessary. If additional space is needed, provide attachments)

The licensee, or any individual executing this certificate on behalf of the licensee, certifies that:

1. No radioactive materials have ever been procured or possessed by the licensee under this license.
2. All activities authorized by this license have ceased, and all radioactive materials procured and/or possessed by the licensee under this license number cited above have been disposed of in the following manner.
- a. Transfer of radioactive materials to the licensee listed below:
RMT, Inc. License #: 025-1246-01 Issued by: State of Wisconsin
- b. Disposal of radioactive materials: Date of transfer 2/15/12 Dustin Huenink 402-429-5317
1. Directly by the licensee:
2. By licensed disposal site:
3. By waste contractor:
- c. All radioactive materials have been removed such that any remaining residual radioactivity is within the limits of 10 CFR Part 20, Subpart E, and is ALARA.

C. SURVEYS PERFORMED AND REPORTED

1. A radiation survey was conducted by the licensee. The survey confirms:
- a. the absence of licensed radioactive materials
- b. that any remaining residual radioactivity is within the limits of 10 CFR 20, Subpart E, and is ALARA.
2. A copy of the radiation survey results:
- a. is attached; or b. is not attached (Provide explanation); or c. was forwarded to NRC on: _____ Date
3. A radiation survey is not required as only sealed sources were ever possessed under this license, and
- a. The results of the latest leak test are attached; and/or b. No leaking sources have ever been identified.

The person to be contacted regarding the information provided on this form:


NAME	TITLE	TELEPHONE (Include Area Code)	E-MAIL ADDRESS
Dustin Huenink	Senior H's mgr. / RSO	402-429-5317	dustin.huenink@pmtinc.com

Mail all future correspondence regarding this license to:

744 Heartland Trail, PO Box 8923, Madison WI 53708

C. CERTIFYING OFFICIAL

I CERTIFY UNDER PENALTY OF PERJURY THAT THE FOREGOING IS TRUE AND CORRECT

PRINTED NAME AND TITLE	SIGNATURE	DATE
Dustin Huenink Senior H's mgr. / RSO		2/16/12

WARNING: FALSE STATEMENTS IN THIS CERTIFICATE MAY BE SUBJECT TO CIVIL AND/OR CRIMINAL PENALTIES. NRC REGULATIONS REQUIRE THAT SUBMISSIONS TO THE NRC BE COMPLETE AND ACCURATE IN ALL MATERIAL RESPECT. 18 U.S.C. SECTION 1001 MAKES IT A CRIMINAL OFFENSE TO MAKE A WILLFULLY FALSE STATEMENT OR REPRESENTATION TO ANY DEPARTMENT OR AGENCY OF THE UNITED STATES AS TO ANY MATTER WITHIN ITS JURISDICTION.



Troxler Electronic Laboratories, Inc.

3008 Cornwallis Rd., P.O. Box 12057
Research Triangle Park, NC 27709
Tel: (877) 876-9537 Fax: (866) 391-2759
License: NC 032-0182-1

DUSTIN HUENINK
RMT, INC.
744 HEARTLAND TRAIL
MADISON, WI 53708-8923

Cust ID: 6726

LEAK TEST CERTIFICATE

DEVICE:

Model: 3440 Serial No: 60347

SEALED SOURCES:

Serial No.	Measure Date	Nuclide	GBq	mCi
77-7437	09/24/2007	Cs-137	0.296	8
47-6272	03/13/1984	Am-241:Be	1.48	40

LEAK TEST ANALYSIS:

Sample collected on: 02/14/2012
Sample analyzed on: 02/15/2012 10:56:08 A Position: 1
Analyzed by: D. Rose

	ALPHA	BETA-GAMMA
Conversion factor (cpm/Bq)	1.25E+01	1.97E+01
Background measurement (cpm)	2	27
Sample measurement (cpm)	1	28
Activity (Bq)	< MDA	< MDA
Min. Detectable Activity (Bq)	6.7E-01	1.4E+00

This certifies that the leak test results are:

Less than 185 Bq (0.005 uCi) Greater than 185 Bq (0.005 uCi)



Troxler Electronic Laboratories, Inc.

3008 Cornwallis Rd., P.O. Box 12057
Research Triangle Park, NC 27709
Tel: (877) 876-9537 Fax: (866) 391-2759
License: NC 032-0182-1

DUSTIN HUENINK
RMT, INC.
744 HEARTLAND TRAIL
MADISON, WI 53708-8923

Cust ID: 6726

LEAK TEST CERTIFICATE

DEVICE:

Model: 3440 Serial No: 60265

SEALED SOURCES:

Serial No.	Measure Date	Nuclide	GBq	mCi
77-7414	09/24/2007	Cs-137	0.296	8
78-4559	09/17/2007	Am-241:Be	1.48	40

LEAK TEST ANALYSIS:

Sample collected on: 02/14/2012
Sample analyzed on: 02/15/2012 10:57:20 A Position: 2
Analyzed by: D. Rose

	ALPHA	BETA-GAMMA
Conversion factor (cpm/Bq)	1.25E+01	1.97E+01
Background measurement (cpm)	2	27
Sample measurement (cpm)	0	29
Activity (Bq)	< MDA	< MDA
Min. Detectable Activity (Bq)	6.7E-01	1.4E+00

This certifies that the leak test results are:

Less than 185 Bq (0.005 uCi) Greater than 185 Bq (0.005 uCi)



Troxler Electronic Laboratories, Inc.

3008 Cornwallis Rd., P.O. Box 12057
Research Triangle Park, NC 27709
Tel: (877) 876-9537 Fax: (866) 391-2759
License: NC 032-0182-1

DUSTIN HUENINK
RMT, INC.
744 HEARTLAND TRAIL
MADISON, WI 53708-8923

Cust ID: 6726

LEAK TEST CERTIFICATE

DEVICE:

Model: 3440 **Serial No:** 39274

SEALED SOURCES:

Serial No.	Measure Date	Nuclide	GBq	mCi
77-6751	04/18/2007	Cs-137	0.296	8
78-4059	05/16/2007	Am-241:Be	1.48	40

LEAK TEST ANALYSIS:

Sample collected on: 02/14/2012
Sample analyzed on: 02/15/2012 10:58:33 A **Position:** 3
Analyzed by: D. Rose

	ALPHA	BETA-GAMMA
Conversion factor (cpm/Bq)	1.25E+01	1.97E+01
Background measurement (cpm)	2	27
Sample measurement (cpm)	0	29
Activity (Bq)	< MDA	< MDA
Min. Detectable Activity (Bq)	6.7E-01	1.4E+00

This certifies that the leak test results are:

Less than 185 Bq (0.005 uCi) **Greater than 185 Bq (0.005 uCi)**



Troxler Electronic Laboratories, Inc.

3008 Cornwallis Rd., P.O. Box 12057
Research Triangle Park, NC 27709
Tel: (877) 876-9537 Fax: (866) 391-2759
License: NC 032-0182-1

DUSTIN HUENINK
RMT, INC.
744 HEARTLAND TRAIL
MADISON, WI 53708-8923

Cust ID: 6726

LEAK TEST CERTIFICATE

DEVICE:

Model: 3440 **Serial No:** 31419

SEALED SOURCES:

Serial No.	Measure Date	Nuclide	GBq	mCi
47-28091	11/20/1998	Am-241:Be	1.48	40
750-6152	01/10/2000	Cs-137	0.296	8

LEAK TEST ANALYSIS:

Sample collected on: 02/14/2012
Sample analyzed on: 02/15/2012 10:59:45 A **Position:** 4
Analyzed by: D. Rose

	ALPHA	BETA-GAMMA
Conversion factor (cpm/Bq)	1.25E+01	1.97E+01
Background measurement (cpm)	2	27
Sample measurement (cpm)	1	21
Activity (Bq)	< MDA	< MDA
Min. Detectable Activity (Bq)	6.7E-01	1.4E+00

This certifies that the leak test results are:

Less than 185 Bq (0.005 uCi) **Greater than 185 Bq (0.005 uCi)**



Troxler Electronic Laboratories, Inc.

3008 Cornwallis Rd., P.O. Box 12057
Research Triangle Park, NC 27709
Tel: (877) 876-9537 Fax: (866) 391-2759
License: NC 032-0182-1

DUSTIN HUENINK
RMT, INC.
744 HEARTLAND TRAIL
MADISON, WI 53708-8923

Cust ID: 6726

LEAK TEST CERTIFICATE

DEVICE:

Model: 3440 **Serial No:** 23165

SEALED SOURCES:

Serial No.	Measure Date	Nuclide	GBq	mCi
47-18992	12/10/1993	Am-241:Be	1.48	40
75-4997	02/10/1994	Cs-137	0.296	8

LEAK TEST ANALYSIS:

Sample collected on: 02/14/2012
Sample analyzed on: 02/15/2012 11:00:57 A **Position:** 5
Analyzed by: D. Rose

	ALPHA	BETA-GAMMA
Conversion factor (cpm/Bq)	1.25E+01	1.97E+01
Background measurement (cpm)	2	27
Sample measurement (cpm)	1	35
Activity (Bq)	< MDA	< MDA
Min. Detectable Activity (Bq)	6.7E-01	1.4E+00

This certifies that the leak test results are:

Less than 185 Bq (0.005 uCi) **Greater than 185 Bq (0.005 uCi)**



Troxler Electronic Laboratories, Inc.

3008 Cornwallis Rd., P.O. Box 12057

Research Triangle Park, NC 27709

Tel: (877) 876-9537 Fax: (866) 391-2759

License: NC 032-0182-1

DUSTIN HUENINK
RMT, INC.
744 HEARTLAND TRAIL
MADISON, WI 53708-8923

Cust ID: 6726

LEAK TEST CERTIFICATE

DEVICE:

Model: 3440 Serial No: 22734

SEALED SOURCES:

Serial No.	Measure Date	Nuclide	GBq	mCi
75-4507	09/01/1993	Cs-137	0.296	8
47-18579	06/29/1993	Am-241:Be	1.48	40

LEAK TEST ANALYSIS:

Sample collected on: 02/14/2012
Sample analyzed on: 02/15/2012 11:02:11 A Position: 6
Analyzed by: D. Rose

	ALPHA	BETA-GAMMA
Conversion factor (cpm/Bq)	1.25E+01	1.97E+01
Background measurement (cpm)	2	27
Sample measurement (cpm)	0	26
Activity (Bq)	< MDA	< MDA
Min. Detectable Activity (Bq)	6.7E-01	1.4E+00

This certifies that the leak test results are:

Less than 185 Bq (0.005 uCi) **Greater than 185 Bq (0.005 uCi)**



Troxler Electronic Laboratories, Inc.

3008 Cornwallis Rd., P.O. Box 12057

Research Triangle Park, NC 27709

Tel: (877) 876-9537 Fax: (866) 391-2759

License: NC 032-0182-1

DUSTIN HUENINK
RMT, INC.
744 HEARTLAND TRAIL
MADISON, WI 53708-8923

Cust ID: 6726

LEAK TEST CERTIFICATE

DEVICE:

Model: 3440 Serial No: 22731

SEALED SOURCES:

Serial No.	Measure Date	Nuclide	GBq	mCi
75-4504	09/01/1993	Cs-137	0.296	8
47-18576	06/29/1993	Am-241:Be	1.48	40

LEAK TEST ANALYSIS:

Sample collected on: 02/14/2012
Sample analyzed on: 02/15/2012 11:03:23 A Position: 7
Analyzed by: D. Rose

	ALPHA	BETA-GAMMA
Conversion factor (cpm/Bq)	1.25E+01	1.97E+01
Background measurement (cpm)	2	27
Sample measurement (cpm)	3	35
Activity (Bq)	< MDA	< MDA
Min. Detectable Activity (Bq)	6.7E-01	1.4E+00

This certifies that the leak test results are:

Less than 185 Bq (0.005 uCi) **Greater than 185 Bq (0.005 uCi)**



Troxler Electronic Laboratories, Inc.

3008 Cornwallis Rd., P.O. Box 12057
Research Triangle Park, NC 27709
Tel: (877) 876-9537 Fax: (866) 391-2759
License: NC 032-0182-1

DUSTIN HUENINK
RMT, INC.
744 HEARTLAND TRAIL
MADISON, WI 53708-8923

Cust ID: 6726

LEAK TEST CERTIFICATE

DEVICE:

Model: 3440 Serial No: 14905

SEALED SOURCES:

Serial No.	Measure Date	Nuclide	GBq	mCi
47-10262	07/21/1987	Am-241:Be	1.48	40
50-3655	07/21/1987	Cs-137	0.296	8

LEAK TEST ANALYSIS:

Sample collected on: 02/14/2012
Sample analyzed on: 02/15/2012 12:15:58 P Position: 8
Analyzed by: D. Rose

	ALPHA	BETA-GAMMA
Conversion factor (cpm/Bq)	1.25E+01	1.97E+01
Background measurement (cpm)	2	27
Sample measurement (cpm)	2	46
Activity (Bq)	< MDA	< MDA
Min. Detectable Activity (Bq)	6.7E-01	1.4E+00

This certifies that the leak test results are:

- Less than 185 Bq (0.005 uCi) Greater than 185 Bq (0.005 uCi)

Huenink
RMT, Inc.
744 Heartland Trail
Madison, WI
53708-8923

5.00.93²
FEB 15 2012
12006-121

First Class Mail

Materials Licensing Section
U.S. Nuclear Regulatory Commission, Region III
2443 Warrenville Rd, Suite 210
Lisle, IL 60532-4352

