

NRC FORM 314
(1-2012)
10 CFR 30.36(j)(1), 40.42(j)(1),
70.38(j)(1), and 72.54(k)(5)(1)(1)

U.S. NUCLEAR REGULATORY COMMISSION

APPROVED BY OMB: NO. 3150-0028

EXPIRES: 10/31/2013

Br. 3

CERTIFICATE OF DISPOSITION OF MATERIALS

Estimated burden per response to comply with this mandatory collection request: 30 minutes. This submittal is used by NRC as part of the basis for its determination that the facility is released for unrestricted use. Send comments regarding burden estimate to the Information Services Branch (1-5 F53), U.S. Nuclear Regulatory Commission, Washington, DC 20555-0001, or by internet e-mail to info@nrc.gov, and to the Desk Officer, Office of Information and Regulatory Affairs, NEOB-10202, (3150-0028), Office of Management and Budget, Washington, DC 20503. If a means used to impose an information collection does not display a currently valid OMB control number, the NRC may not conduct or sponsor, and a person is not required to respond to, the information collection.

LICENSEE NAME AND ADDRESS

S&R Engineering S.E.
15 Lepanto Street
Rio Piedras, Puerto Rico
P.O. Box 29406
65TH Infantry Station
San Juan, Puerto Rico 00929

LICENSE NUMBER

52-30913-01

DOCKET NUMBER

030-36569

LICENSE EXPIRATION DATE

June 30, 2014

A. LICENSE STATUS (Check the appropriate box)

- This license has expired.
- This license has not yet expired; please terminate it.

B. DISPOSAL OF RADIOACTIVE MATERIAL

(Check the appropriate boxes and complete as necessary. If additional space is needed, provide attachments)

The licensee, or any individual executing this certificate on behalf of the licensee, certifies that:

- 1. No radioactive materials have ever been procured or possessed by the licensee under this license.
- 2. All activities authorized by this license have ceased, and all radioactive materials procured and/or possessed by the licensee under this license number cited above have been disposed of in the following manner.
 - a. Transfer of radioactive materials to the licensee listed below:
Earth Engineers, Marcos Arocho, Lic. # 52-31427-01, Docket # 030-38405
 - b. Disposal of radioactive materials:
 - 1. Directly by the licensee:
 - 2. By licensed disposal site:
 - 3. By waste contractor:
 - c. All radioactive materials have been removed such that any remaining residual radioactivity is within the limits of 10 CFR Part 20, Subpart E, and is ALARA.

C. SURVEYS PERFORMED AND REPORTED

- 1. A radiation survey was conducted by the licensee. The survey confirms:
 - a. the absence of licensed radioactive materials
 - b. that any remaining residual radioactivity is within the limits of 10 CFR 20, Subpart E, and is ALARA.
- 2. A copy of the radiation survey results:
 - a. is attached; or b. is not attached (Provide explanation); or c. was forwarded to NRC on: 02-10-2012
Date
- 3. A radiation survey is not required as only sealed sources were ever possessed under this license, and
 - a. The results of the latest leak test are attached; and/or
 - b. No leaking sources have ever been identified.


The person to be contacted regarding the information provided on this form:

NAME	TITLE	TELEPHONE (Include Area Code)	E-MAIL ADDRESS
Jaime Sánchez Rivera	President	787-282-8224	jsr@saber.com

Mail all future correspondence regarding this license to:
P.O. Box 29406 Rio, 65TH Infantry Station, San Juan, Puerto Rico 00929

C. CERTIFYING OFFICIAL

I CERTIFY UNDER PENALTY OF PERJURY THAT THE FOREGOING IS TRUE AND CORRECT

PRINTED NAME AND TITLE	SIGNATURE	DATE
Jaime Sánchez Rivera, President		02/10/2012

WARNING: FALSE STATEMENTS IN THIS CERTIFICATE MAY BE SUBJECT TO CIVIL AND/OR CRIMINAL PENALTIES. NRC REGULATIONS REQUIRE THAT SUBMISSIONS TO THE NRC BE COMPLETE AND ACCURATE IN ALL MATERIAL RESPECT. 18 U.S.C. SECTION 1001 MAKES IT A CRIMINAL OFFENSE TO MAKE A WILLFULLY FALSE STATEMENT OR REPRESENTATION TO ANY DEPARTMENT OR AGENCY OF THE UNITED STATES AS TO ANY MATTER WITHIN ITS JURISDICTION.

Leak Test Record
NRC License 62-25430-01

Leak Tested For: S & R Engineering
Leak Tested By: David Rhoe

Standard Source: Cs-137 NES-139S
Standard Activity: 0.105 uCi Nominally
Date of Standard: 32395

Date of Leak Test: 3-Feb-12
Decay Activity: 0.08115 From decay chart
Standard (dpm): 135753.0

Instrument: Beckman Gamma
Instrument Model Number: 5500
Instrument Serial Number: 8044768

Standard (cpm) 35508
Efficiency: 0.28166 28.2 %
Counting time (minutes) 1
Background (cpm) 86
Minimum Detectable Activity: 2.258E-05

Wipe (Smear) Test: All external or accessible surfaces of the source or housing are wiped with a piece of filter paper or other absorbent material which has been moistened with an appropriate solvent and the activity removed is measured.

	Gamma	
Source ID and Serial Number	Wipe Test	Sample Activity
Am-241 & Cs-137 SnM340707502	107	0.00018

This test reveals that 0.005 microcuries or less was present as removable contamination. Should the removable contamination exceed 0.005 microcuries, the source must be removed from use and necessary measures taken according to NRC regulations.


David Rhoe Health/Medical Physicist



February 14, 2012.

To: Mr. William M. Dean
Regional Administrator

Re: Nuclear Gauge Docket No. 03036569, License No. 52-30913-01

Dear Mr. Dean:

By this means we inform you of the transfer agreement between S&R Engineering and Earth Engineers, represented by their corresponding presidents, Mr. Jaime Sánchez Rivera and Mr. Marcos O. Arocho Ramírez respectively, and establish the following:

That Earth Engineers will pay for this transfer the amount \$ 1000.00.

That weekly payment of \$ 200.00 will be done by Earth Engineers until full payment is completed.

That S&R Engineering will transfer the Nuclear Gauge once Earth Engineers receive proper authorization from NRC.

That Earth Engineers has been authorized By NRC under License No. 52-31427-01, Docket No. 030-38405 to receive this Nuclear Gauge from S&R Engineers.


That Earth Engineers has taken the Nuclear Gauge to an authorized inspector to perform an inspection and leak survey at his own cost.


That the Nuclear Gauge is in the facilities of David Rhoe, Health/Medical Physicist, where an inspection and leak survey has been performed.

That with the sign of this document Earth Engineers is in complete possession of the Nuclear Gauge.

That with the sign of this document, S&R Engineers gives complete ownership of the Nuclear Gauge to Earth Engineers.

We here certify this agreement:


Jaime Sánchez Rivera
President
S&R Engineers, S.F.


Marcos O. Arocho Ramírez
President
Earth Engineers



February 14, 2012.

To: Mr. William M. Dean
Regional Administrator

Re: Nuclear Gauge Docket No. 03036569, License No. 52-30913-01

Dear Mr. Dean:

As requested in your letter of January 13 of 2012, specifically in page 5, point 2b, regarding the last time the Nuclear Gauge was used, we can confirm that was during the summer of 2008 in a project located in the San German area.

If you need additional information, please let me know.

Regards,

A handwritten signature in black ink, appearing to read 'Jaime Sanchez Rivera', is written over the typed name.

Jaime Sanchez Rivera
President
S&R Engineering, S.E.

This is to acknowledge the receipt of your letter/application dated

2/10/2012 and to inform you that the initial processing which includes an administrative review has been performed.

Termination (52-30913-01)
There were no administrative omissions. Your application was assigned to a technical reviewer. Please note that the technical review may identify additional omissions or require additional information.

Please provide to this office within 30 days of your receipt of this card

A copy of your action has been forwarded to our License Fee & Accounts Receivable Branch, who will contact you separately if there is a fee issue involved.

Your action has been assigned **Mail Control Number** 576994.
When calling to inquire about this action, please refer to this control number.
You may call us on (610) 337-5398, or 337-5260.