

U.S. NUCLEAR REGULATORY COMMISSION

Date: 09/18/12

TELEPHONE CONVERSATION RECORD

Time: 09:43am and 10:30am

Mail Control
or Report No(s).

577809

License No(s). 06-08349-04

Docket No(s). 03009293

Name of Licensee: Charlotte Hungerford Hospital

Name of Participant(s): Jennifer Tyrian, Manager of Diagnostic Services
Gerald J. Randall, Radiation Safety Officer

Telephone No. (860) 496-6552
(860) 559-9366

Subject: Additional information for amendment request

(NOTE: This will be used as the Documents Title in ADAMS)

NRC staff contacted the licensee to clarify the following information provided in an amendment request letter dated June 15, 2012:

1. I informed Ms. Tyrian that in order to complete transfer, Charlotte Hungerford Hospital has to either accept the facilities at Connecticut Oncology & Hematology as is or Connecticut Oncology & Hematology would need to provide close out surveys of their facilities. Ms. Tyrian stated that Charlotte Hungerford Hospital will be accepting the facilities as is, and agreed to provide documentation of the commitment by COB today.
2. Shielding of the facility at Connecticut Oncology & Hematology described in the letter received November 3, 2010 (Mail Control 573305). Mr. Randall confirmed that the shielding originally designed and described (including the PET scanning room, quiet rooms, and hot lab, and adjacent areas), remains the same as that provided in the letter and there are no plans to make any changes to the current facility. He also stated that he performed surveys around the facility and found it to be adequate.
3. F-18 for calibration. The licensee asked to remove the user, Sadek Nehmeh, Ph.D., for this use, from their license. Mr. Randall confirmed that they would like to retain F-18 for calibration for future use, and requested to have him, Gerald J. Randall, listed as the user.

Action Required: Issue Amendment**Document Availability:** Publicly Available Non-Publicly Available
 Non-Sensitive
 Sensitive – Privacy Act/ PII
 Immediate Release

 Non-Sensitive Copyright
 Sensitive – Internal
 Normal Release
 Date:09/26/2012

 Sensitive- Proprietary
 Sensitive – Security-Related
 Delayed Release Date:

SUNSI Review Completed By: M. Abogunde / RA /

Document Accession #: