

U.S. NUCLEAR REGULATORY COMMISSION

Date: 09/19/12

TELEPHONE CONVERSATION RECORD

Time: 10:00am

Mail Control
or Report No(s).

577810

License No(s). 06-31418-01

Docket No(s). 03038340

Name of Licensee: Connecticut Oncology & Hematology
Name of Participant(s): Cathy Coleman, Nuclear Medicine Technician
Telephone No. (860) 482-8691
Subject: Additional information for termination request
 (NOTE: This will be used as the Documents Title in ADAMS)

NRC staff contacted the licensee to clarify the following information provided in a deficiency response dated September 12, 2012:

1. The Cs-137 source (Model GF-0008, SN 1181-63-82) listed on the leak test and the inventory records showed different activities (0.5 mCi and 0.5 μ Ci). Ms. Coleman verified that the actual activity of the Cs-137 rod source is 0.5 nanocuries (nCi), and that the units reflected on the other records resulted from auto fill typos when the documents were created.

Action Required: Issue termination**Document Availability:****Publicly Available****Non-Publicly Available****Non-Sensitive****Non-Sensitive Copyright****Sensitive- Proprietary****Sensitive – Privacy Act/ PII****Sensitive – Internal****Sensitive – Security-Related****Immediate Release****Normal Release****Delayed Release Date:**

Date:09/27/2012

SUNSI Review Completed By:

M. Abogunde

/ RA /

Document Accession #: