

September 14, 2012

Mr. Blake Welling
U.S. Nuclear Regulatory Commission
Region I
Materials Security and Industrial Branch
Division of Nuclear Materials Safety
2100 Renaissance Blvd, Suite 100
King of Prussia, PA 19406-2713

Re: Response to Inspection and Notice of Violations (Revised).

Dear Mr. Welling:

The following specific corrective actions have been taken in response to your inspection of June 25, 2012 and the Notice of Violation of July 2, 2012. Mr. Ivan Usero-Perez is the new RSO for Geopractica.

Violation 1. Physical inventory - Compliance was achieved on August 13, 2012. It will be performed every 6 months. Also, the audit of the program was undertaken.

Reason for the violation – Resignation of the RSO and transition time in which Mr. Ivan Usero Perez was able to comply with responsibilities.

Corrective steps – The physical inventory was conducted. Note reports dated August 13, 2012 attached to our letter of August 13, 2012. Mr. Ivan Usero-Perez will take action to avoid further violations with the assistance of our consultant in NRC matters, Mr. David Rhoe.

Violation 2. Leak tests - Compliance was achieved on August 4, 2012. They will be tested every 6 months.

Reason for the violation – Resignation of the RSO and transition time in which Mr. Ivan Usero Perez was able to comply with responsibilities.

Corrective steps – The gauges were tested for leakage on August 4, 2012. Note reports dated August 4, 2012 attached to our letter of August 13, 2012. Mr. Ivan Usero-Perez will take action to avoid further violations with the assistance of our consultant in NRC matters, Mr. David Rhoe.

Violation 3. HAZMAT training course - Compliance was achieved on August 4, 2012. The course will be taken every 3 years.

Reason for the violation – Resignation of the RSO and transition time in which Mr. Ivan Usero Perez was able to comply with responsibilities.

Corrective steps – The HAZMAT training course was taken on August 4, 2012. Note certificates dated August 4, 2012 attached to our letter of August 13, 2012. Mr. Ivan Usero-Perez will take action to avoid further violations with the assistance of our consultant in NRC matters, Mr. David Rhoe.

Therefore, full compliance was achieved on August 13, 2012. Please take note that the survey meter was calibrated, the notice to employees was posted, the certificate of competent authority for sealed sources is available and the shipping papers have been updated.

Respectfully,

Ivan Usero Perez

President

Geopractica, Inc.