2001 Rankin Road Houston, Texas 77373 Tel (713) 879-2662 Fax (713) 879-2667 Cell (713) 306-0658

Philip G. Simpkin Radiation Safety Officer phil.simpkin@bakerhughes.com



Baker Hughes Oilfield Operations, Inc.

28th March 2014

Ms. Michelle M Hammond M.Sc., Senior Health Physicist United States Regulator Commission Region IV 1600 East Lamar Blvd., Arlington, Texas 76011



DNMS

42-19649-01

Re: Office covered by RAM License No. 42-029649-01

Dear Ms. Hammond,

I would like to inform you that Baker Hughes is no longer storing radioactive materials at its Gaylord Michigan facility.

The facility is listed on the license at 1250 Anna Drive, Gaylord Michigan

The sources and records from the facility have all been transferred to the one in Mill Hall Pennsylvania.

Baker Hughes has no plans to store radioactive materials at the Gaylord facility and would like to remove it from the license.

Should you have any questions or require and additional information, please feel free to contact me

Sincerely,

Philip G. Simpkin CRadP Radiation Safety Officer,

Baker Hughes Oilfield Operations, Inc.

PUBLIC

Immediate Release

Normal Release

NON-PUBLIC

☐ A.3 Sensitive-Security Related

☐ A.7 Sensitive Internal

☐ Other:

Reviewer:

Date: 2 1

Gaylord Michigan Close out Survey


Dylan Dailey, RSO

August 12th, 2014



Background

Baker Hughes is requesting closure





Surveys taken

- The Radiation Safety Officer for Baker Hughes(Dylan Dailey) performed a closeout survey of the facility on August14, 2014.
- The perimeter of the facility and the interior surfaces of each building was thoroughly scanned using a RadEye GN+. No reading of greater than 10 µRem was noted anywhere on the facility.
- In addition, the entire facility was walked on a 5 foot grid using a Thermo-Electron Interceptor in the finder mode to ensure no orphan sources of radiation were left behind.
- Leak Tests have been reviewed for all sources stored at this facility, and no leaking sources were found.
- 3 Area Wipe tests were taken on the man doors entering the workshop and office area. No removable contamination above background Beta/Gamma was observed. No Alpha disintegrations were observed. (see attachment).





PUBLIC

Immediate Release
Normal Release

NON-PUBLIC

- ☐ A.3 Sensitive-Security Related☐ A.7 Sensitive Internal
- C Other:

Reviewer:



Work Shop



Yard

SUNTRAC SERVICES, INC. 1818 EAST MAIN STREET LEAGUE CITY, TEXAS 77573 PHONE (281) 338-2133

AREA WIPE TEST REPORT FORM

Company Name: Baker Hughes-Pressure Pumping

Smear Taken Date: August 14, 2014 Smears Taken By: Dylan Dailey

This report certifies that wipe test smears are from <u>Baker Hughes-Pressure Pumping</u> located at <u>1245 Anna Dr. Gaylord, MI 49735</u> and they have been assayed at our facility for removable contamination with a Ludlum 2929, S/N 86481 Calibration certificate attached.

Smear Location	Alpha DPM	Beta/Gamma DPM	
#1 - main ETEC lab	BDL	BDL	
#2 -down ETEC lab	BDL	BDL	
#3 – up ETEC lab	BDL	BDL	
#4 - source store gate	BDL	BDL	
#5 – office doors	BDL	BDL	

BDL = Below Detectable Limits

CERTIFIED BY:

VIVE JUWIN

August 20, 2014

QC'D BY:

DATE:

August 20 2014



SUNTRAC Services, Inc. CERTIFICATE OF CALIBRATION

1818 East Main Street League City, TX 77573 Phone: (281) 338-2133 Fax: (281) 338-2136 www.suntrac.com

Customer: 147

Suntrac - IN HOUSE METERS

1818 East Main Street League City, TX 77573

		Manufact	urer	Model Numb	oer	Serial Number
Instrument Serviced:		Ludlum		2929 (L)		86481
	Detector:	Ludlum		43-10-1		PR108772
Meter Range Multiplier	Distance Calculated (inches)	Filter	Calculated Calibration Point	Instrument Meter Reading	Unit	Correction Factor
Alpha Channel	Pulser	N	40	41	C.P.M.	1.0
Alpha Channel	Pulser	N	400	400	C.P.M.	
Alpha Channel	Pulser	N	4000	3999	C.P.M.	
Alpha Channel	Pulser	N	40000	40033	C.P.M.	
Alpha Channel	Pulser	N	400000	401101	C.P.M.	
Beta/Gamma Cł	Pulser	N	40	40	C.P.M.	1.0
Beta/Gamma Cl	Pulser	N	400	401	C.P.M.	
Beta/Gamma Cl	Pulser	N	4000	4010	C.P.M.	
Beta/Gamma Cl	Pulser	N	40000	40076	C.P.M.	
Beta/Gamma Cl	Pulser	N	400000	401253	C.P.M.	
Minutes	Pulser	N	0.1	.1	Min	1.0
Minutes	Pulser	N	1	1	Min	
Minutes	Pulser	N	10	10	Min	

Calibrated in accordance with Texas Radioactive Materials License Number L03062, and requirements of ANSI/NCSL Z540-1-1994; ANSI-N323A-1997; and 10 CFR 34.25 & 35.61. NIST Traceable #111926/T111433.

Temperature: 79°F

Humidity: 38%

Altitude: 30inHG

Geotropism: ✓

Calibration Source S/N: A280

Source Activity: 58.40 mCi

Cs137: 20.10 mR/hr.@100cm;<+/-5%

Pulser Utilized: Yes

Pulser Information: Ludlum Model 500 S/N 70634

Battery Check: N/A

OP Volts: 825

Detector Position: Parallel

Check Source: No

Isotope:

Isotope S/N:

Min At Contact:

Technician Comments: This instrument is calibrated electronically in C. P. M.

Keith lager

7/16/14

6 months

1/16/15

Calibrated By Keith Sager

Cal. Date

Calibration Interval Cal. Due Date



SUNTRAC Services, Inc. CERTIFICATE OF CALIBRATION

1818 East Main Street League City, TX 77573 Phone: (281) 338-2133 Fax: (281) 338-2136 www.suntrac.com

Customer: 147

Suntrac - IN HOUSE METERS

1818 East Main Street League City, TX 77573

		Manufacturer		Model Number		Serial Number	
Instrument Serviced:		Ludlum		2929 (L)		86481	
	Detector:	Ludlum		43-10-1		PR108772	
Meter Range Multiplier	Distance Calculated (inches)	Filter	Calculated Calibration Point	Instrument Meter Reading	Unit	Correction Factor	
HV	Pulser	N	500	500	Volts	1.0	
HV	Pulser	N	2000	2000	Volts		

Calibrated in accordance with Texas Radioactive Materials License Number L03062, and requirements of ANSI/NCSL Z540-1-1994; ANSI-N323A-1997; and 10 CFR 34.25 & 35.61. NIST Traceable #111926/T111433.

Temperature: 79°F

Humidity: 38%

Altitude: 30inHG

Geotropism: ✓

Calibration Source S/N: A280

30

Source Activity: 58.40 mCi

Cs137: 20.10 mR/hr.@100cm;<+/-5%

Pulser Utilized: Yes

Pulser Information: Ludlum Model 500 S/N 70634

Battery Check: N/A

OP Volts: 825

Detector Position: Parallel

Check Source: No

Isotope:

Isotope S/N:

Volts At Contact:

Technician Comments: This instrument is calibrated electronically in C. P. M.

PUBLIC

Immediate Release
Normal Release

NON-PUBLIC

A.3 Sensitive S.

A.7 Sensitive internal

1/16/15

Calibrated By Keith Sager

Keith lager

7/16/14 Cal. Date

Calibration Interval Cal. Due Date

This certificate may not be reproduced except in full, without written permission from Suntrac Services

BETWEEN: Accounts Receivable/	Payable	[FOR ARPB USE] INFORMATION FROM WBL		
and Regional Licensing Br		Program Code: 03120 Status Code: Pending Amendment Fee Category: 3P Exp. Date: Fee Comments: 3P EFF 7/28/89 Decom Fin Assur Reqd: N		
License Fee Wo	orksheet - License Fe	ee Transmittal		
A. REGION				
1. APPLICATION ATTAC Applicant/Licensee: Received Date: Docket Number: Mail Control Number: License Number: Action Type:		ns, Inc. Through Its Pressure Pumping Service Line		
2. FEE ATTACHED Amount:				
Check No.:				
3. COMMENTS				
	Signed: Colle	en Murnahan		
	Date: 9-3	2019		
	EMENT BRANCH (Check when	milestone 03 is entered / /)		
Fee Category and Am	ount:	-		
2. Correct Fee Paid. Appl	ication may be processed for:			
Amendment:				
Renewal:				
License:				
3. OTHER				
	Signed:			

Date:

NRC FORM 532 (1-2012)



DATE

Falson Annual Comments of the	09/16/2014				
NAME AND ADDRESS OF APPLICANT AND/OR LICENSEE	LICENSE NUMBER				
Baker Hughes Oilfield Operations, Inc. ATTN: Philip G. Simpkin CRadP	42-19649-01				
	MAIL CONTROL NUMBER				
Radiation Safety Officer 2001 Rankin Road	584724				
Houston, TX 77074	LICENSING AND/OR TECHNICAL REVIEWER				
	cmurnahan UM				
This is to acknowledge the receipt of your:					
✓ LETTER and/or ✓ APPL	ICATION DATED: 08/27/2014				
The initial processing, which included an adn	ninistrative review, has been performed.				
✓ AMENDMENT TERMINATIO	✓ AMENDMENT TERMINATION NEW LICENSE RENEWAL				
✓ There were no administrative omissions ide	There were no administrative omissions identified during our initial review.				
This is to acknowledge receipt of your application for renewal of the material(s) license identified above. Your application is deemed timely filed, and accordingly, the license will not expire until final action has been taken by this office.					
	Your application for a new NRC license did not include your taxpayer identification number. Please fill out NRC Form 531, located at the following link:				
http://www.nrc.gov/readir	ng-rm/doc-collections/forms/nrc531.pdf				
Send the completed NRC Form 531, by fac	ssimile, to the following number: (301) 415-5387				
	our License Fee and Accounts Receivable Branch, in ou will be contacted separately if there is a fee issue				
calling to inquire about this action, please rebeen forwarded to a technical reviewer. Plenormally completed within 180 days for a re	ove listed MAIL CONTROL NUMBER. When efer to this control number. Your application has ease note that the technical review, which is enewal application (90 days for all other requests), additional information. If you have any questions on, our contact information is listed below:				
Region IV U. S. Nuclear Regulator	ry Commission				

DNMS/NMSB - B 1600 E. Lamar Blvd. Arlington, TX 76011-4511 (817) 200-1103 or (817) 200-1140

emailed to become