



UNITED STATES
NUCLEAR REGULATORY COMMISSION
 REGION IV
 1600 EAST LAMAR BLVD
 ARLINGTON, TEXAS 76011-4511

EMAIL



Name: Steve Roberts License: 11-27073-01
Organization: The Amalgamated Sugar Company, LLC Docket: 030-32226
Phone: 208-438-7176 Control: 608910
E-mail Address: SCRoberts@amalsugar.com
From: Jacqueline D. Cook
Date: August 9, 2018
Subject: Email dated May 2, 2018 and Supplemental Information dated July 24, 2018 and August 6, 2018 for License Amendment
Pages: 2

Mr. Roberts:

Per your email dated May 2, 2018, and supplemental information dated July 24, 2018 and August 6, 2018, the item on the next page is a deficiency which require your response. **Please respond to this e-mail as soon as practical but no later than Thursday, August 23, 2018.** Our fax number is (817) 200-1263. **Please provide a response in a signed and dated letter in pdf format when responding via email.** If additional time is needed to submit your response, please do not hesitate to contact me with an alternative response date. My email address is Jackie.Cook@nrc.gov. When responding to this e-mail, please include the license, docket, and control numbers located at the top of this page.

Thanking you in advance for your cooperation, assistance, and prompt response in this matter.

/RA/
 Jacqueline D. Cook
 Senior Health Physicist

PUBLIC

- Immediate Release
- Normal Release

NON-PUBLIC

- A.3 Sensitive-Security Related
- A.7 Sensitive Internal
- Other:

Reviewer: JDC Date: _____

1. As I was preparing to issue the amendment to your license, Items 7.A. through 7.D. caught my attention so I looked up the manufacturer and model of the devices in the Sealed Source & Device Registry (SS&DR) (i.e., Berthold Technologies GmbH & Co. KG [Berthold Technologies USA, LLC] Models LB 7400 Series: LB 7440D, LB 7440F, and LB 7442F [TN-1031-D-101-B]; Laboratorium Prof. Dr. Berthold [Berthold Systems, Inc.] Model LB 300 IRL Series [Drawing 21257.000-000] [NR-8050-D-807-S]) and will plan to update Item 7 accordingly based on these devices being registered and evaluated through the SS&DR. Please see attached draft license.

Please note that I was unable to find the Tracor, Inc. Model 550 GC in the SS&DR. Please confirm the correct manufacturer and model of the nickel-63 source.

Please provide an inventory of your radioactive material you have or update my attached draft license including the correct manufacturer and model of the nickel-63 source so your license may be amended accordingly and appropriately. Please update Item 8.A. through 8.E. of the license also specifying the possession limit per source and the maximum possess at any one time.